

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130114</u> ✓
Name <u>Kevin Stevenson</u>		ID# <u>2414</u> ✓
Date <u>9/1/10</u>		<u>WAA PD</u>
A	Agency <u>WAA Police Dept</u>	Phone # <u>786-1120</u>
Instrument Location <u>WAA Police Dept suite 114 ESH</u>		
B	Alco S/N <u>75470</u> ✓	Target Value <u>.080</u> ✓
		High Pressure <u>200</u>
Alco Test Values		
	<u>.081</u> ✓	<u>.081</u> ✓
	1 st Alco	2 nd Alco
Signature <u>Kevin Stevenson</u>		
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 4th day of October, 2010.

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130114

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

Breath Test Record			
Serial Number	130114	Model	DataMaster
Location	ANC7	University Police - UAA	
Subject Information			
Last Name	Verification	O.L. #	12345
First Name	Verification		
Operator Information			
Operator Name	Stevenson Kevin L	Case	102394
Operator #	2414	Department	ANC7
Test Results			
Alco Target	0.080	Alco Serial	75470
Nominal	0.079	Barometer	29.72
Dry Gas 1	0.081	Time	00:11:33
Subject Alcohol	0.000	Time	00:13:01
Dry Gas 2	0.081	Time	00:14:06
		Test Type	V
		Test Date	09/01/2010
		Test Time	00:10:09
		Report	Y

Diagnostic Test Record	
Serial Number	130114
Model	DataMaster
Location	ANC7
University Police - UAA	
Diagnostic Test Results	
Test Date	09/01/2010
Test Time	00:07:13
Computer	Okay
Program	Okay
Software Date	02/20/2001
Sample Chamber (C)	48
Breath Tube (C)	41
Barometer (in)	29.72
Flow Detector	Okay
High Speed Pump	Okay
Detector	Okay
Filter	Okay
Quartz	Okay
Calibration	Okay

--- BREATH ANALYSIS ---

ALCO ADJUSTED FOR 29.72 in
ALCO TARGET 0.080
BLANK TEST 0.000
INTERNAL STANDARD 0.000
ALCO TO 29.72 in 0.080
BLANK TEST 0.000
SUBJECT SAMPLE 0.000
BLANK TEST 0.000
ALCO TO 29.72 in 0.080
BLANK TEST 0.000

8MB
9/3/10

PUMP SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

9/3/2010 00:10:09
DataMaster S/N 130114
Operator: Stevenson, Kevin L
Location: ANC7